



4800 MALDI and AB SCIEX 5800 MALDI™ TOF/TOF MASS SPECTROMETER

CE DECLARATION OF CONFORMANCE

5008184A

↑__

↓_

↓__



DECLARATION OF CONFORMITY

(According to EN ISO/IEC 17050-1:2004)



We

AB SCIEX

of

71 Four Valley Drive, Concord, ON, Canada L4K 4V8

declare under sole responsibility that the product as originally delivered

Equipment: **TOF/TOF Mass Spectrometer**
Product Name: **4800 MALDI and AB SCIEX 5800 MALDI™**
Product Options: **All**

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2006/95/EC - Low Voltage Directive**
- 2004/108/EC - Electromagnetic Compatibility Directive**
- 2002/96/EC - Waste Electrical and Electronic Equipment Directive**

and conforms with the following product standards:

Category	Standard	Classification
Safety	IEC 61010-1:2001 / EN 61010-1:2001	Equipment Class I, Pollution Degree 2
	IEC 61010-2-061:2001 / EN 61010-2-061:2003	
	Canada: CAN/CSA-C22.2 No. 61010-1-04	
	CAN/CSA-C22.2 No. 61010-2-061-04	
USA:	ANSI/UL 61010-1:2004, IEC 61010-2-061:2001	
EMC	IEC 61326-1:2005/ EN 61326-1:2006	Limit Class A Group 1 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 0.5 kV signal lines, 1 kV AC lines 2 kV common mode, 0.5 kV differential mode 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short interruption: 0% for 250 cycles. Class A Group 1 Class A Group 1 Class A
	Reference Standards	
	CISPR 11 / EN 55011	
	IEC 61000-4-2 / EN 61000-4-2	
	IEC 61000-4-3 / EN 61000-4-3	
	IEC 61000-4-4 / EN 61000-4-4	
	IEC 61000-4-5 / EN 61000-4-5	
	IEC 61000-4-6 / EN 61000-4-6	
	IEC 61000-4-8 / EN 61000-4-8	
	IEC 61000-4-11 / EN 61000-4-11	
	Australia/New Zealand: AS/NZS CISPR 11:2004	
	Canada: CISPR 11:2004	
	USA: FCC Part 15	

The product was tested in a typical configuration.

Year Mark First Applied: **2005**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

Gordon Logan

Senior Vice President, Research and Development

Keith Price

Senior Vice President, Quality and Regulatory

2010-07-19

Date

This product was developed and structurally validated following a Registered Quality Management system, which conforms to ISO 9001 International Standard Requirements, as audited and registered by the British Standards Institution under ISO 9001 Certificate Number: FM 31547

